

LOT: 11704-52
EXP: 2017-09-25
Open Vial Stability: 16 days

Medonic CA530, CA620		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8,6	0,6	8,0	9,2
Hct	%	29,2	2,5	26,7	31,7
MCV	fL	74,6	5,0	69,6	79,6
RBC	10 ¹² /L	3,91	0,18	3,73	4,09
Hgb	g/dL	10,9	0,4	10,5	11,3
MCH	pg	27,9	1,5	26,4	29,4
MCHC	g/dL	37,3	2,2	35,1	39,5
PLT	10 ⁹ /L	229	30	199	259

EXIGO		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8,0	0,6	7,4	8,6
Hct	%	30,0	2,5	27,5	32,5
MCV	fL	77,7	5,0	72,7	82,7
RBC	10 ¹² /L	3,86	0,18	3,68	4,04
Hgb	g/dL	11,0	0,4	10,6	11,4
MCH	pg	28,5	1,5	27,0	30,0
MCHC	g/dL	36,7	2,2	34,5	38,9
PLT	10 ⁹ /L	189	30	159	219

This assay sheet is also available on :

www.medonic.se
www.exigo-vet.com

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 11704-52
Prod.: Vet-Con Norm
ID: 1170452+
Exp.: 25-Sep-2017

