

LOT: 11707-52
EXP: 2017-12-25
Open Vial Stability: 16 days

Medonic CA530, CA620		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8,5	0,6	7,9	9,1
Hct	%	31,2	2,5	28,7	33,7
MCV	fL	78,6	5,0	73,6	83,6
RBC	10 ¹² /L	3,97	0,18	3,79	4,15
Hgb	g/dL	11,6	0,4	11,2	12,0
MCH	pg	29,2	1,5	27,7	30,7
MCHC	g/dL	37,2	2,2	35,0	39,4
PLT	10 ⁹ /L	215	30	185	245

EXIGO		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8,1	0,6	7,5	8,7
Hct	%	32,0	2,5	29,5	34,5
MCV	fL	81,4	5,0	76,4	86,4
RBC	10 ¹² /L	3,93	0,18	3,75	4,11
Hgb	g/dL	11,6	0,4	11,2	12,0
MCH	pg	29,5	1,5	28,0	31,0
MCHC	g/dL	36,3	2,2	34,1	38,5
PLT	10 ⁹ /L	186	30	156	216

This assay sheet is also available on :

www.medonic.se
www.exigo-vet.com

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 11707-52
Prod.: Vet-Con Norm
ID: 1170752+
Exp.: 25-Dec-2017

