

LOT: 11804-52
EXP: 2018-09-24
Open Vial Stability: 16 days

Medonic CA530, CA620		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8,7	0,6	8,1	9,3
Hct	%	30,0	2,5	27,5	32,5
MCV	fL	77,3	5,0	72,3	82,3
RBC	10 ¹² /L	3,88	0,18	3,70	4,06
Hgb	g/dL	11,1	0,4	10,7	11,5
MCH	pg	28,6	1,5	27,1	30,1
MCHC	g/dL	37,0	2,2	34,8	39,2
PLT	10 ⁹ /L	237	30	207	267

EXIGO		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8,0	0,6	7,4	8,6
Hct	%	31,1	2,5	28,6	33,6
MCV	fL	80,4	5,0	75,4	85,4
RBC	10 ¹² /L	3,87	0,18	3,69	4,05
Hgb	g/dL	11,2	0,4	10,8	11,6
MCH	pg	28,9	1,5	27,4	30,4
MCHC	g/dL	36,0	2,2	33,8	38,2
PLT	10 ⁹ /L	210	30	180	240

This assay sheet is also available on :

www.medonic.se
www.exigo-vet.com

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 11804-52
Prod.: Vet-Con Norm
ID: 1180452+
Exp.: 24-Sep-2018

