

This assay sheet is also available on :

www.medonic.sewww.exigo-vet.com

LOT: 11610-52
EXP: 2017-03-28
Open Vial Stability: 16 days

Medonic CA530, CA620		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	8.2	0.6	7.6	8.8
Hct	%	31.1	2.5	28.6	33.6
MCV	fL	76.9	5.0	71.9	81.9
RBC	10 ¹² /L	4.04	0.18	3.86	4.22
Hgb	g/dL	11.9	0.4	11.5	12.3
MCH	pg	29.5	1.5	28.0	31.0
MCHC	g/dL	38.3	2.2	36.1	40.5
PLT	10 ⁹ /L	231	30	201	261

EXIGO		Mean	± Range	Expected Range	
WBC	10 ⁹ /L	7.8	0.6	7.2	8.4
Hct	%	32.0	2.5	29.5	34.5
MCV	fL	79.9	5.0	74.9	84.9
RBC	10 ¹² /L	4.01	0.18	3.83	4.19
Hgb	g/dL	11.9	0.4	11.5	12.3
MCH	pg	29.7	1.5	28.2	31.2
MCHC	g/dL	37.2	2.2	35.0	39.4
PLT	10 ⁹ /L	204	30	174	234

Manufactured for:
Boule Medical AB
Domnarvsgatan 4
SE-163 53 Spånga, Sweden



1504056

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"

2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 11610-52
Prod.: Vet-Con Norm
ID: 1161052+
Exp.: 28-Mar-2017

